Zafety Lug Lock[™] Torque Test Report June 25, 2009

A series of tests were carried out on lug nuts with Zafety Lug LocksTM installed to verify the torqued readings.

The test was completed under the following conditions:

- 1. Onsite at a company that has been using Zafety Lug LocksTM on their trailers.
- 2. Test equipment (certification document on separate page):

Manufacturer: Torqueleader

Model: DDS800S Serial No.: 0GZ036048

Asset No.:M1220

Tool Capacity: 600 lbf.ft
Accuracy of set torque <u>+</u> 4%
Calibration Date: June 24, 2009

- 3. Technician performing the tests: Topline Trailer Service Ltd.
- 4. Test Location: Toronto Ontario

Test 1:

Steel Rims, Uni-Wheel

Rim Installation Date: November 2008

Torque applied: 500 lbf.ft

Zafety Lug LocksTM Installation Date: November 2008

Test date: June 24, 2009

Removal Torque Applied: 500 lbf.ft

Results:

No evidence of movement.

Test 2:

Alloy Rims, Uni-Wheel

Rim Installation Date: November 2008

Torque applied: 500 lbf.ft

Zafety Lug LocksTM Installation Date: November 2008

Test date: June 24, 2009

Removal Torque Applied: 500 lbf.ft

Results:

No evidence of movement.

Attachments:

Calibration Certification'
Onsite Photos

Customer Name.: TAFCAN CONSULTING LTD.

.. TAI CAR CORCOLING EID

2107 PINEVALLEY CRESCENT, OAKVILLE, ONTARIO

CERTIFICATE OF CALIBRATION

McCANN EQUIPMENT LTD.



2178 Torquay Mews Mississauga, ON.

www.torquetools.com

Accredited to ISOIIEC 17025:2005 Certificate Number: L2097.02-1

Fax: (905) 542 1335

341747

LABORATORY

ACCREDITATION

01

Tel: (9

Manufacturer:

Tool Capacity:

Model:

CW

Address.:

(905) 542 1333

1 800 287 5714

Torqueleader

DDS800S

600

4%

ueleader

0GZ036048

Cert No.:

Issue No.:

Asset No.:

M1220

Unit of Measure:

lbf.ft

Ambient Temperature:

24 °C

Humidity:

Serial No.:

49%

Procedure No.:

MCC 6.00

Accuracy of set torque ± : Technician:

Quality Supervisor:

David Mackaren

Kathy Quart KO

CCW

120.00 360.00

600.00

Lower Limit	Upper Limit
115.20	124.80
345.60	374.40
576.00	624.00

As Found	As Left
117.86	117.86
365.05	365.05
611.64	611.64

Reading are derived from the average of five readings taken at the "set torque" as indicated Statistical information is available upon request.

Transfer Standards

Display Model: 43228

Transducer: 50012ETS

Serial No.: 60241

Serial No.: 34021ETS

STATEMENT OF COMPLIANCE:

McCann Equipment Limited hereby certifies that the above described instrument has been calibrated in compliance with the requirements and specifications in the Purchase Order, using measurement equipment and standards that are traceable to the NRC/CANADA and / or the NIST/USA.

McCann Equipment Limited Quality System meets or exceeds the requirements of ISO 17025.

The estimated uncertainty of measurement is ± 0.48% of applied load, K=2, 95% confidence level.

The Test Equipment used for this calibration complies with the requirements of ISO 10012.

The limits shown for this calibration comply, or exceed the requirements of ISO 6789.

This certificate is not to be considered as a "guarantee of accuracy" for the indicated Calibration Interval.

This certificate shall not be reproduced, except in full, without written approval of McCann Equipment Limited.

Received Date:	24-Jun-09
Calibration Date:	24-Jun-09
Calibration Due Date:	24-Jun-10

Condition	Out of Spec.	In Spec.
Received		X
Returned		X

Remarks:

Before and after readings remain the same.

For information, please contact:

Kathy Quart









